

Quick Start Guide

Cordless Companion Monitor Standalone Kit

Applies to product codes:

R09-00011 | R09-0003 | R09-0005 | R09-0007 | R09-0007U |
R09-00091 | R09-0011 | R09-0013 | R09-0015

System Contents	
Monitor	RG2 Cordless Companion Monitor
Sensor	RG11 / RG49 Cordless Seatgard RG13 / RG47 Cordless Bedgard RG59 Cordless Floorgard RG39 Cordless Under-mattress Pad
Transmitter	RG9 Cordless Pad Transmitter

Sensor Placement

RG13 / RG47 Bedgard: Place on top of the mattress and under the bed sheet. The pad can be placed anywhere between the shoulders to the hips depending on the condition of the patient and how early you want the alert.

RG39 Under-mattress: Place under the mattress across the width of the bed at the lower back level. On profiling beds make sure it will not entrap the mechanism when the bed is being adjusted.

For further guidance on placement, please refer to the pad positioning flyer.

RG11 / RG49 Seatgard: Place mat on the relevant seat on top of or under a cushion, giving consideration to the position of the monitor and the least intrusive method of installing the equipment.

RG59 Floorgard: Place mat at or near to the area to be monitored such as a doorway or side of the bed.

Receiver/Monitor Placement

The monitor can be positioned on a wall via the plastic wall bracket or on a bed or chair via the hanging strap.

General Operation

The monitor triggers an alarm when the sensor mat is activated and can be reset at the monitor or by reactivating the pad.



Quick Start Guide

Cordless Companion Monitor Standalone Kit

Follow the steps below to set up your equipment:

- 1.** Pull out the battery tab to power up the monitor or insert 3 AA batteries into the monitor. A long beep will confirm the batteries have been inserted correctly.
- 2.** Fit the rubber cover and mount the monitor in the desired location using the supplied hanging strap or wall bracket.
- 3.** Ensure that the sensor pad is placed on a smooth flat surface and the pad is not creased or folded in any way.
- 4.** Check that you have positioned your sensor pad correctly and switch the transmitter on (inside pouch).
- 5.** To test, activate the sensor to confirm that it is alarming and resetting correctly.

When you first switch on the pad transmitter, apply pressure to the sensor pad for at least 15 secs to wake up the transmitter from sleep mode. Failure to do this will cause the Companion to loose connection from the transmitter and trigger an audible alert. This only applies when you switch a transmitter from OFF to ON for the first time.

When you are satisfied with the points above please complete the daily checks sheet provided before putting the system into use.

**Ramblegard Technical Support Helpline:
028 9448 3979**

